



# FCC DoC TEST REPORT

for

**Medical Station**

**MODEL: xxxxxONYX-172DTy-xxxxxxx; xxxxxONYX-192DTy-xxxxxxx**

Test Report Number:  
90323201-D

Issued to:

**AAEON Technology Inc.**

**5F, No.135, Lane 235, Pao Chiao Rd., Hsin-Tien City,  
Taipei, Taiwan, R.O.C.**

Issued by:

**Compliance Certification Services Inc.**

**Sindian BU.**

**No.163-1, Jhongsheng Rd., Sindian City,  
Taipei County 23151, Taiwan**

**TEL: 886-2-22170894**

**FAX: 886-2-22171029**

**Issued Date: April 13, 2009**



***Note:** This report shall not be reproduced except in full, without the written approval of Compliance Certification Services Inc. This document may be altered or revised by Compliance Certification Services Inc. personnel only, and shall be noted in the revision section of the document. The client should not use it to claim product endorsement by TAF, A2LA, NVLAP, NIST or any government agencies. The test results in the report only apply to the tested sample.*



**Revision History**

Rev.	Issue Date	Revisions	Effect Page	Revised By
00		Initial Issue	ALL	



**TABLE OF CONTENTS**

**1 TEST RESULT CERTIFICATION..... 4**

**2 EUT DESCRIPTION..... 5**

**3 TEST METHODOLOGY ..... 6**

3.1. DECISION OF FINAL TEST MODE..... 6

3.2. EUT SYSTEM OPERATION ..... 6

**4 SETUP OF EQUIPMENT UNDER TEST ..... 7**

4.1. DESCRIPTION OF SUPPORT UNITS ..... 7

4.2. CONFIGURATION OF SYSTEM UNDER TEST..... 8

**5 FACILITIES AND ACCREDITATIONS ..... 9**

5.1. FACILITIES ..... 9

5.2. ACCREDITATIONS ..... 9

5.3. MEASUREMENT UNCERTAINTY ..... 9

**6 CONDUCTED EMISSION MEASUREMENT..... 10**

6.1. LIMITS OF CONDUCTED EMISSION MEASUREMENT ..... 10

6.2. TEST INSTRUMENTS ..... 10

6.3. TEST PROCEDURES ..... 11

6.4. TEST SETUP ..... 12

6.5. DATA SAMPLE..... 12

6.6. TEST RESULTS..... 13

**7 RADIATED EMISSION MEASUREMENT ..... 14**

7.1. LIMITS OF RADIATED EMISSION MEASUREMENT ..... 14

7.2. TEST INSTRUMENTS ..... 14

7.3. TEST PROCEDURES ..... 15

7.4. TEST SETUP ..... 16

7.5. DATA SAMPLE..... 17

7.6. TEST RESULTS..... 18

**8 PHOTOGRAPHS OF THE TEST CONFIGURATION ..... 20**



# 1 TEST RESULT CERTIFICATION

**Product:** Medical Station

**Brand:** AAEON

**Model:** xxxxxONYX-172DTy-xxxxxxx (Where y is T or blank and x is 0-9,A-Z,-or blank);  
xxxxxONYX-192DTy-xxxxxxx (Where y is T or blank and x is 0-9,A-Z,-or blank)

**Applicant:** **AAEON Technology Inc.**  
5F, No.135, Lane 235, Pao Chiao Rd., Hsin-Tien City,  
Taipei, Taiwan, R.O.C.

**Manufacturer:** **AAEON Technology Inc.**  
5F, No.135, Lane 235, Pao Chiao Rd., Hsin-Tien City,  
Taipei, Taiwan, R.O.C.

**Tested:** March 23, 2009 & March 24, 2009

Applicable Standard	Class / Limit	Test Result
FCC Part 18	---	No non-compliance noted
<b>Deviation from Applicable Standard</b>		
None		

*Note: 1. The statements of test result on the above are decided by the request of test standard only; the measurement uncertainties are not factored into this compliance determination.  
2. The information of measurement uncertainty is available upon the customer's request.*

Deviation from Applicable Standard
None

The above equipment has been tested by Compliance Certification Services Inc., and found compliance with the requirements set forth in the technical standards mentioned above. The results of testing in this report apply only to the product/system, which was tested. Other similar equipment will not necessarily produce the same results due to production tolerance and measurement uncertainties.

**Approved by:**

**Reviewed by:**

Vince Chiang  
Assistant Manager of Sindian BU.

Vesta Hsu  
Supervisor of report document dept. of Sindian BU.



## 2 EUT DESCRIPTION

<b>Product</b>	Medical Station
<b>Trade Name</b>	AAEON
<b>Model</b>	xxxxxONYX-172DTy-xxxxxxx (Where y is T or blank and x is 0-9,A-Z,-or blank); xxxxxONYX-192DTy-xxxxxxx (Where y is T or blank and x is 0-9,A-Z,-or blank)
<b>Applicant</b>	AAEON Technology Inc.
<b>Housing Type</b>	Plastic w/ metal plate
<b>Identify Number</b>	90323201
<b>Received Date</b>	March 23, 2009
<b>EUT Power Rating</b>	9~30VDC from AC Adaptor
<b>AC Power During Test</b>	120VAC / 60 Hz to AC Adaptor
<b>AC Adaptor Manufacturer</b>	PROTEK POWER
<b>AC Adaptor Model Number</b>	PMP120-14-B16
<b>Power Adaptor Power Rating</b>	I/P: 100~240VAC; O/P: 24V/5A
<b>AC Power Cable Type</b>	Unshielded, 1.8m (Detachable) to Power Adaptor
<b>DC Power Cable Type</b>	Unshielded, 1.2m (Non-detachable, with a core) to AC Adaptor
<b>OSC/Clock Frequencies</b>	25MHz; 24.576MHz; 14.31818MHz; 32.768kHz

### Model Differences:

	Model Name	Difference	Tested (Checked)
<b>Original</b>	ONYX-172DTT	17"	<input checked="" type="checkbox"/>
<b>Additional</b>	ONYX-192DTT	19"	<input checked="" type="checkbox"/>

### I/O PORT

I/O PORT TYPES	Q'TY	TESTED WITH
1) SIO Port	2	2
2) VGA Port	1	1
3) USB Port	2	2
4) LAN Port	1	1
5) PS/2 one to two adaptor	1	1

*Note: None.*

### 3 TEST METHODOLOGY

#### 3.1. DECISION OF FINAL TEST MODE

The EUT was tested together with the above additional components, and a configuration, which produced the worst emission levels, was selected and recorded in this report.

The test configuration/ modes is as the following:

**Conduction Modes:**

1	ONYX-172DTT	D-SUB ONLY MODE (1920X1440, VF-60Hz)
2		D-SUB ONLY MODE (1600X1200, VF-60Hz)
3		EUT+D-SUB MODE (1280X1024, VF-60Hz)
4		EUT+D-SUB MODE (1024X768, VF-60Hz)
5		EUT+D-SUB MODE (800X600, VF-60Hz)
6	ONYX-192DTT	EUT+D-SUB MODE (1024X768, VF-60Hz)

**Radiation Modes:**

1	ONYX-172DTT	D-SUB ONLY MODE (1920X1440, VF-60Hz)
2		D-SUB ONLY MODE (1600X1200, VF-60Hz)
3		EUT+D-SUB MODE (1280X1024, VF-60Hz)
4		EUT+D-SUB MODE (1024X768, VF-60Hz)
5		EUT+D-SUB MODE (800X600, VF-60Hz)
6	ONYX-192DTT	EUT+D-SUB MODE (1280X1024, VF-60Hz)
		EUT+D-SUB MODE (1280X1024, VF-60Hz)/ 1-7.5GHz

**Conduction:** Mode 6

**Radiation:** Mode 6

#### 3.2. EUT SYSTEM OPERATION

1. Windows XP boots system.
2. Run Emctest.exe to activate all peripherals and display “H” pattern on monitor screen.
3. Run Winemc.exe and choose “F:/ & G:/ & H:/” to test USB 2.0 Ports.
4. Press the start menu, select executive and type ping 192.168.0.1 –t (EUT), ping 192.168.0.2 –t (Server PC).

*Note: Test program is self-repeating throughout the test.*



## 4 SETUP OF EQUIPMENT UNDER TEST

### 4.1. DESCRIPTION OF SUPPORT UNITS

The EUT has been tested as an independent unit together with other necessary accessories or support units. The following support units or accessories were used to form a representative test configuration during the tests.

#### EUT Devices:

No.	Equipment	Model No.	Trade Name
1	CPU	Celeron M Processor 1.5GHz	Intel
2	PANEL (TFT LCD 17")	CLAA170EA07 / 4 LAMP	CPT (For model name: ONYX-172DTT)
	PANEL (TFT LCD 19")	HSD190MEN3-A00 / 4 CCFLS Backlight System	HannStar (For model name: ONYX-192DTT)
3	Hard Disk	MHW2080AT/80G	Fujitsu
4	Memory (1GB DDR 333MHz)	SAMSUNG K4H51083BD-ZCCC	Transcend


#### Peripherals Devices:

No.	Equipment	Model No.	Serial No.	FCC ID / BSMI ID	Trade Name	Data Cable	Power Cord
1-2	USB 2.0 HDD X2	F12-U	N/A	BSMI ID: 4912A002	TeraSys	Shielded, 1.8m	N/A
3	PS/2 Mouse	M071KC	443029438	DOC BSMI: R41108	DELL	Shielded, 1.8m	N/A
4	PS/2 Keyboard	SK-8110	N/A	DOC BSMI: T3A002	DELL	Shielded, 1.8m	N/A
5-6	Modem X2	5JEG4033MKO	N/A	5RJTAI-35500-M5-E	TOP-SOLUTION	Unshielded, 1.8m	Unshielded, 1.8m
7	Monitor	202P40	BZ000405640107	FCC ID: A3KM107 BSMI: R33048	PHILIPS	Shielded, 1.8m with two cores	Unshielded, 1.8m
8	Server PC	dc7100 CMT	SGH43200NP	DOC BSMI: R33001	HP	Unshielded, 10m	Unshielded, 1.8m

#### Note:

- 1) All the equipment/cables were placed in the worst-case configuration to maximize the emission during the test.
- 2) Grounding was established in accordance with the manufacturer's requirements and conditions for the intended use.

#### 4.2. CONFIGURATION OF SYSTEM UNDER TEST





## 5 FACILITIES AND ACCREDITATIONS

### 5.1. FACILITIES

All measurement facilities used to collect the measurement data are located at CCS Taiwan Sindian BU. at No.163-1, Jhongsheng Rd., Sindian City, Taipei County 23151, Taiwan.

The sites are constructed in conformance with the requirements of ANSI C63.4 and CISPR Publication 22. All receiving equipment conforms to CISPR 16-1-1, CISPR 16-1-2, CISPR 16-1-3, CISPR 16-1-4 and CISPR 16-1-5.

### 5.2. ACCREDITATIONS

Our laboratories are accredited and approved by the following accreditation body according to ISO/IEC 17025.

<b>Taiwan</b>	TAF
<b>USA</b>	A2LA

The measuring facility of laboratories has been authorized or registered by the following approval agencies.

<b>Canada</b>	Industry Canada
<b>Germany</b>	TUV Rheinland
<b>Japan</b>	VCCI
<b>Taiwan</b>	BSMI
<b>USA</b>	FCC

Copies of granted accreditation certificates are available for downloading from our web site, <http://www.ccsemc.com>

### 5.3. MEASUREMENT UNCERTAINTY

Where relevant, the following measurement uncertainty levels have been estimated for tests performed on the EUT as specified in CISPR 16-4-2:

Measurement	Frequency	Uncertainty
Conducted emissions	0.15MHz~30MHz	± 1.7376
Radiated emissions	30MHz ~ 200MHz	± 3.9041
	200MHz ~1000MHz	± 3.9162

This uncertainty represents an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of k=2.

Consistent with industry standard (e.g. CISPR 22: 2006, clause 11, Measurement Uncertainty) determining compliance with the limits shall be base on the results of the compliance measurement. Consequently the measure emissions being less than the maximum allowed emission result in this be a compliant test or passing test.

The acceptable measurement uncertainty value without requiring revision of the compliance statement is base on conducted and radiated emissions being less than  $U_{CISPR}$  which is 3.6dB and 5.2dB respectively. CCS values (called  $U_{Lab}$  in CISPR 16-4-2) is less than  $U_{CISPR}$  as shown in the table above. Therefore, MU need not be considered for compliance.



## 6 CONDUCTED EMISSION MEASUREMENT

### 6.1. LIMITS OF CONDUCTED EMISSION MEASUREMENT

#### Maximum permissible level of Line Conducted Emission

Frequency (MHZ)	(dBuV)	
	Quasi-peak	Average
0.15-0.5	66-56	56-46
0.50 - 5.0	56	46
5.0 - 30.0	60	50

**NOTE:**

- (1) The lower limit shall apply at the transition frequencies.
- (2) The limit decreases in line with the logarithm of the frequency in the range of 0.15 to 0.50 MHz.
- (3) All emanations from a class A/B digital device or system, including any network of conductors and apparatus connected thereto, shall not exceed the level of field strengths specified above.

### 6.2. TEST INSTRUMENTS

Conducted Emission Test Site # A				
Name of Equipment	Manufacturer	Model	Serial Number	Calibration Due
TEST RECEIVER	R&S	ESHS20	840455/006	02/12/2010
LISN (EUT)	SCHWARZBECK	NSLK 8127	8127382	12/09/2009
LISN	SOLAR	8012-50-R-24-BNC	8305114	12/09/2009
BNC CABLE	MIYAZAKI	5D-FB	BNC A4	05/12/2009
THERMO-HYGRO METER	TECPEL	DTM-303	No.7	11/24/2009
Test S/W	EMI 32.exe			

- NOTE:** 1. The calibration interval of the above test instruments is 12 months and the calibrations are traceable to NML/ROC and NIST/USA.  
 2. N.C.R = No Calibration Request.



### **6.3. TEST PROCEDURES** (please refer to measurement standard or CCS SOP PA-031)


#### **Procedure of Preliminary Test**

- The EUT and Support equipment, if needed, was set up as per the test configuration to simulate typical usage per the user's manual. When the EUT is a tabletop system, a wooden table with a height of 0.8 meters is used and is placed on the ground plane as per ANSI C63.4 (see Test Facility for the dimensions of the ground plane used). When the EUT is a floor standing equipment, it is placed on the ground plane, which has a 12 mm non-conductive covering to insulate the EUT from the ground plane.
- All I/O cables were positioned to simulate typical actual usage as per ANSI C63.4.
- The test equipment EUT installed received AC main power, through a Line Impedance Stabilization Network (LISN), which supplied power source and was grounded to the ground plane.
- All support equipment power received from a second LISN.
- The EUT test program was started. Emissions were measured on each current carrying line of the EUT using an EMI Test Receiver connected to the LISN powering the EUT.
- The Receiver scanned from 150kHz to 30MHz for emissions in each of the test modes.
- During the above scans, the emissions were maximized by cable manipulation.
- The test mode(s) described in Item 3.1 were scanned during the preliminary test.
- After the preliminary scan, we found the test mode described in Item 3.1 producing the highest emission level.
- The EUT configuration and cable configuration of the above highest emission levels were recorded for reference of the final test.

#### **Procedure of Final Test**

- EUT and support equipment were set up on the test bench as per the configuration with highest emission level in the preliminary test.
- A scan was taken on both power lines, Line 1 and Line 2, recording at least the six highest emissions. Emission frequency and amplitude were recorded into a computer in which correction factors were used to calculate the emission level and compare reading to the applicable limit.
- The test data of the worst-case condition(s) was recorded.

**6.4. TEST SETUP**



- For the actual test configuration, please refer to the related item – Photographs of the Test Configuration.

**6.5. DATA SAMPLE**

Freq. MHz	Read Level dBuV	Factor dB	Level dBuV	Limit dBuV	Over Limit dB	Remark (P/Q/A)	Line (L1/L2)
x.xx	42.95	0.55	43.50	56	-12.50	Q	L1

- Freq. = Emission frequency in MHz
- Read Level = Uncorrected Analyzer/Receiver reading
- Factor = Insertion loss of LISN + Cable Loss
- Level = Read Level + Factor
- Limit = Limit stated in standard
- Over Limit = Reading in reference to limit
- P = Peak Reading
- Q = Quasi-peak Reading
- A = Average Reading
- L1 = Hot side
- L2 = Neutral side

**Calculation Formula**

Over Limit (dB) = Level (dBuV) – Limit (dBuV)



6.6. TEST RESULTS

<b>Model No.</b>	ONYX-192DTT	<b>6dB Bandwidth</b>	10 KHz
<b>Environmental Conditions</b>	24deg.C, 58% RH, 1010hPa	<b>Test Mode</b>	Mode 6
<b>Tested by</b>	John Yen		

(The chart below shows the highest readings taken from the final data.)

Six Highest Conducted Emission Readings							
Frequency Range Investigated				150 KHz to 30 MHz			
Freq. (MHz)	Read Level (dBuV)	Factor (dB)	Level (dBuV)	Limit Line (dBuV)	Over Limit (dB)	Remark (P/Q/A)	Line (L1/L2)
0.435	43.07	0.08	43.15	57.15	-14.00	P	L1
0.604	40.89	0.09	40.98	56.00	-15.02	P	L1
0.775	39.31	0.10	39.41	56.00	-16.59	P	L1
0.435	42.85	0.09	42.94	57.15	-14.21	P	L2
0.598	40.43	0.10	40.53	56.00	-15.47	P	L2
1.810	37.01	0.14	37.15	56.00	-18.85	P	L2

**NOTE:** 1. 0.15MHz to 30MHz test is Applicable Part 15.107 standard.  
 2. The emission level was or more than 2dB below the Average limit, so no re-check anymore.



### 7 RADIATED EMISSION MEASUREMENT

#### 7.1. LIMITS OF RADIATED EMISSION MEASUREMENT

**Maximum permissible level of Radiated Emission measured at 3meter**

Frequency (MHZ)	Maximum Field Strength Limit (dBu V/m/ Q.P.)
30-88	40
88-216	43.5
216-960	46
Above 960	53.9

*Note: The lower limit shall apply at the transition frequency.*

Frequency (MHz)	dBuV/m (At 3m)	
	Average	Peak
Above 960	54	74

**NOTE:** (1) The lower limit shall apply at the transition frequencies.  
(2) Emission level (dBuV/m) = 20 log Emission level (uV/m).

#### 7.2. TEST INSTRUMENTS

Open Area Test Site # I				
Name of Equipment	Manufacturer	Model	Serial Number	Calibration Due
MEASURE RECEIVER	SCHAFFNER	SCR3501	338	07/07/2009
SPECTRUM ANALYZER	ADVANTEST	R3132	120900008	No Calibration Required
ANTENNA	SCHAFFNER	CBL 6112B	2809	09/08/2009
AMPLIFIER	SCHAFFNER	CPA9231A	3626	10/12/2009
CABLE	BELDEN	9913	N-TYPE #12	02/22/2010
THERMO-HYGRO METER	TECPEL	DTM-303	080268	05/11/2009
Test S/W	Lab VIEW 7.1			
Above 1GHz Used				
SPECTRUM ANALYZER (9kHz-30GHz)	R&S	FSP 30	100112	10/16/2009
ANTENNA (1-18GHz)	EMCO	3115	00022256	01/22/2010
AMPLIFIER (1-18GHz)	HP	8449B	3008A01266	01/19/2010
CABLE (1-18GHz)	JYEBAO	LL142	SMA#RS1	01/19/2010
CABLE (1-18GHz)	HUBER +SUHNER	SUCOFLEX 104	SMA#RS3	01/19/2010
CABLE (1-18GHz)	JYEBAO	LL142	SMA#C1	01/19/2010
Test S/W	EZ-EMC			

**NOTE:** 1. The calibration interval of the above test instruments is 12 months and the calibrations are traceable to NML/ROC and NIST/USA.  
2. N.C.R = No Calibration Request.



### **7.3. TEST PROCEDURES** (please refer to measurement standard or CCS SOP PA-031)


#### **Procedure of Preliminary Test**

- The equipment was set up as per the test configuration to simulate typical usage per the user's manual. When the EUT is a tabletop system, a wooden turntable with a height of 0.8 meters is used which is placed on the ground plane. When the EUT is a floor standing equipment, it is placed on the ground plane which has a 12 mm non-conductive covering to insulate the EUT from the ground plane.
- Support equipment, if needed, was placed as per ANSI C63.4.
- All I/O cables were positioned to simulate typical usage as per ANSI C63.4.
- The EUT received AC power source from the outlet socket under the turntable. All support equipment power received from another socket under the turntable.
- The antenna was placed at 10/3 meter away from the EUT as stated in ANSI C63.4. The antenna connected to the Spectrum Analyzer via a cable and at times a pre-amplifier would be used.
- The Analyzer / Receiver quickly scanned from 30MHz to 7500MHz. The EUT test program was started. Emissions were scanned and measured rotating the EUT to 360 degrees and positioning the antenna 1 to 4 meters above the ground plane, in both the vertical and the horizontal polarization, to maximize the emission reading level.
- The test mode(s) described in Item 3.1 were scanned during the preliminary test:
- After the preliminary scan, we found the test mode described in Item 3.1 producing the highest emission level.
- The EUT and cable configuration, antenna position, polarization and turntable position of the above highest emission level were recorded for the final test.

**Procedure of Final Test**

- EUT and support equipment were set up on the turntable as per the configuration with highest emission level in the preliminary test.
- The Analyzer / Receiver scanned from 30MHz to 7500MHz. Emissions were scanned and measured rotating the EUT to 360 degrees, varying cable placement and positioning the antenna 1 to 4 meters above the ground plane, in both the vertical and the horizontal polarization, to maximize the emission reading level.
- Recorded at least the six highest emissions. Emission frequency, amplitude, antenna position, polarization and turntable position were recorded into a computer in which correction factors were used to calculate the emission level and compare reading to the applicable limit and only Q.P. reading is presented.
- The test data of the worst-case condition(s) was recorded.

**7.4. TEST SETUP**



- For the actual test configuration, please refer to the related item - Photographs of the Test Configuration.





7.5. DATA SAMPLE

Below 1GHz

Freq. (MHz)	Reading (dBuV/m)	Factor (dB)	Result (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector (P/Q/A)	Pol. (H/V)
x.xx	14.0	12.2	26.2	30	-3.8	Q	H

- Freq. = Emission frequency in MHz
- Reading = Uncorrected Analyzer/Receiver reading
- Factor = Antenna Factor + Cable Loss + Attenuator (3/6/10dB) – Amplifier Gain
- Amptd = Uncorrected Analyzer/Receiver reading + Factor
- Limit = Limit stated in standard
- Margin = Reading in reference to limit
- P = Peak Reading
- Q = Quasi-peak Reading
- A = Average Reading
- H = Antenna Polarization: Horizontal
- V = Antenna Polarization: Vertical

Calculation Formula

Margin (dB) = Result (dBuV/m) – Limit (dBuV/m)

Above 1GHz

Freq. (MHz)	Reading (dBuV/m)	Factor (dB)	Result (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector (P/A)	Pol. (H/V)
x.xx	42.95	0.55	43.50	54	-10.50	A	H

- Freq. = Emission frequency in MHz
- Reading = Uncorrected Analyzer/Receiver reading
- Factor = Antenna Factor + Cable Loss - Amplifier Gain
- Result = Reading + Factor
- Limit = Limit stated in standard
- Margin = Result – Limit
- P = Peak Reading
- A = Average Reading
- H = Antenna Polarization: Horizontal
- V = Antenna Polarization: Vertical



7.6. TEST RESULTS

Below 1GHz

<b>Model No.</b>	ONYX-192DTT	<b>Test Mode</b>	Mode 6
<b>Environmental Conditions</b>	25deg.C, 80% RH, 1010hPa	<b>6dB Bandwidth</b>	120 KHz
<b>Antenna Pole</b>	Vertical / Horizontal	<b>Antenna Distance</b>	3 m
<b>Detector Function</b>	Quasi-peak.	<b>Tested by</b>	John Yen

(The chart below shows the highest readings taken from the final data.)

Six Highest Radiated Emission Readings							
Frequency Range Investigated				30 MHz to 1000 MHz at 3m			
Freq. (MHz)	Reading (dBuV/m)	Factor (dB)	Result (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector (P/Q/A)	Pol. (H/V)
36.803	41.80	-13.92	27.88	40.00	-12.12	Q	V
153.902	43.26	-16.47	26.79	43.50	-16.71	Q	V
200.023	45.30	-17.03	28.27	43.50	-15.23	Q	V
300.010	44.10	-12.57	31.53	46.00	-14.47	Q	V
500.020	39.00	-7.43	31.57	46.00	-14.43	Q	V
756.025	33.10	-3.49	29.61	46.00	-16.39	Q	V

Six Highest Radiated Emission Readings							
Frequency Range Investigated				30 MHz to 1000 MHz at 3m			
Freq. (MHz)	Reading (dBuV/m)	Factor (dB)	Result (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector (P/Q/A)	Pol. (H/V)
137.710	43.61	-15.52	28.09	43.50	-15.41	Q	H
200.178	45.12	-17.03	28.09	43.50	-15.41	Q	H
250.008	49.00	-13.97	35.03	46.00	-10.97	Q	H
324.015	47.20	-11.93	35.27	46.00	-10.73	Q	H
356.071	46.60	-11.04	35.56	46.00	-10.44	Q	H
933.009	35.15	-1.36	33.79	46.00	-12.21	Q	H

- REMARKS:**
1. 30MHz to 1000MHz test is Applicable Part 15.109 standard.
  2. The other emission levels were very low against the limit.
  3. P= Peak Reading; Q= Quasi-peak Reading.



Above 1GHz

<b>Model No.</b>	ONYX-192DTT	<b>Test Mode</b>	Mode 6
<b>Environmental Conditions</b>	26deg.C, 60% RH, 1010hPa	<b>6dB Bandwidth</b>	1000 KHz
<b>Antenna Pole</b>	Vertical / Horizontal	<b>Antenna Distance</b>	3m
<b>Detector Function</b>	Peak oraverage.	<b>Tested by</b>	John Yen

(The chart below shows the highest readings taken from the final data.)

Highest Radiated Emission Readings							
Frequency Range Investigated				30 MHz to 7500 MHz at 3m			
Freq. (MHz)	Reading (dBuV)	Factor (dB)	Result (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Remark (P/Q/A)	Pol. (H/V)
1238.000	55.52	-10.31	45.21	74.00	-28.79	P	V
1882.000	50.62	-6.63	43.99	74.00	-30.01	P	V
2456.000	58.34	-4.44	53.90	74.00	-20.10	P	V
2806.000	51.49	-3.01	48.48	74.00	-25.52	P	V
3408.000	53.22	-1.06	52.16	74.00	-21.84	P	V
3898.000	50.89	0.55	51.44	74.00	-22.56	P	V

Highest Radiated Emission Readings							
Frequency Range Investigated				30 MHz to 7500 MHz at 3m			
Freq. (MHz)	Reading (dBuV)	Factor (dB)	Result (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Remark (P/Q/A)	Pol. (H/V)
1504.000	52.72	-8.98	43.74	74.00	-30.26	P	H
1798.000	52.17	-7.15	45.02	74.00	-28.98	P	H
2218.000	50.97	-5.20	45.77	74.00	-28.23	P	H
2456.000	56.94	-4.44	52.50	74.00	-21.50	P	H
3100.000	52.85	-1.92	50.93	74.00	-23.07	P	H
3450.000	50.67	-0.94	49.73	74.00	-24.27	P	H
3842.000	52.77	0.36	53.13	74.00	-20.87	P	H

**NOTE:** 1. The other emission levels were very low against the limit.  
 2. P= Peak Reading; A= Average Reading.

**8 PHOTOGRAPHS OF THE TEST CONFIGURATION**  
**LINE CONDUCTED EMISSION TEST**



**RADIATED EMISSION TEST**

