



DEGREES OF PROTECTION OF Medical PC ENCLOSURE TEST REPORT

Report No. : IPT20070066

Issue Date : MARCH 23, 2007

Laboratory Name : Lighting Laboratory

Laboratory Address : No. 6-6, Ronggong S. Rd.,
Caota Village, Guanyin Township, Taoyuan County
328, TAIWAN, R.O.C.

Laboratory Registration No. : 1126

大電力



財團法人台灣大電力研究試驗中心
Taiwan Electric Research & Testing Center

地址：328 桃園縣觀音鄉草漯村榮工南路 6-6 號
電話：(03) 483-9090 (代表號)
傳真：(03) 483-8119 (代表號)
電子信箱：customer_service@ms.tertec.org.tw
網址：www.tertec.org.tw

- ◆ This report applies only to the sample.
- ◆ This report is invalid without seal code.
- ◆ Without written permission, this report is not allowed to be duplicated partially, but it does not limit to complete copy.
- ◆ This report is not allowed to be used as advertisement, publication, and sales promotion material.
- ◆ Inquiry telephone : +886-3-483-9090



財團法人

台灣大電力研究試驗中心
Taiwan Electric Research & Testing Center

Lighting Laboratory

Report No. : IPT20070066

DEGREES OF PROTECTION OF Medical PC ENCLOSURE

Applicant : AAEON TECHNOLOGY INC.

Applicant Address : F5, NO.135, LANE 235, PAO CHIAO RD., HSIN-TIEN CITY, TAIPEI,
TAIWAN, R.O.C.

Manufacturer : AAEON TECHNOLOGY INC.

Manufacturer Address : F5, NO.135, LANE 235, PAO CHIAO RD., HSIN-TIEN CITY, TAIPEI,
TAIWAN, R.O.C

Sample : Medical PC

Model / Brand Name : ONYX-153 / —

Dimensions : 398 mm (L) ×330 mm (W) ×92.5 mm (D)

Standard : IEC 60529 (2001-02)

Sampling procedure : Sent by applicant.

Sampling date : MARCH 20, 2007

Testing date : MARCH 23, 2007

Testing engineer : Cheng, Sheng-Hsing

Result of test : The enclosure of Medical PC (Model : ONYX-153) is satisfactory for IPX1.

Result of test refers to page 2.

Signatory of the report. 



Result of Test

1.Result of test: The enclosure of Medical PC is satisfactory for IPX1.

2.Description of test:

Test items	Standards of test	Result of test
Degrees of protection against water. Second characteristic numeral 1.	IEC60529 (2001-02) sec.14.1 sec.14.2.1 sec.14.3	The protection is satisfactory.

1



財團法人

台灣大電力研究試驗中心


Taiwan Electric Research & Testing Center

Lighting Laboratory


Report No. : IPT20070066

3. Pictures of sample:

Front View



Rear View





財團法人

台灣大電力研究試驗中心

Taiwan Electric Research & Testing Center

Lighting Laboratory

Report No. : IPT20070066

Internal View

