



Lighting Laboratory
1126



DEGREES OF PROTECTION OF Medical PC ENCLOSURE TEST REPORT

Report No. : IPT20090089-011

Issue Date : August 10, 2009

Laboratory Name : Lighting Laboratory

Laboratory Address : No. 6-6, Ronggong S. Rd.,

Caota Village, Guanyin Township, Taoyuan County
328, TAIWAN, R.O.C.

Laboratory Registration No. : 1126

大電力



財團法人台灣大電力研究試驗中心
Taiwan Electric Research & Testing Center

ADD : 328 桃園縣觀音鄉草漯村榮工南路 6-6 號

TEL : (03) 483-9090 (Generation of style-name)

FAX : (03) 483-8119 (Generation of style-name)

MAIL : customer_service@ms.tertec.org.tw

WEBSITE : www.tertec.org.tw

- ◆ This report applies only to the sample.
- ◆ This report is invalid without seal code.
- ◆ Without written permission, this report is not allowed to be duplicated partially, but it does not limit to complete copy.
- ◆ This report is not allowed to be used as advertisement, publication, and sales promotion material.
- ◆ Inquiry telephone : +886-3-483-9090



財團法人

台灣大電力研究試驗中心

Taiwan Electric Research & Testing Center

Lighting Laboratory

Report No. : IPT20090089-011

DEGREES OF PROTECTION OF Medical PC ENCLOSURE

Applicant : AAEON Technology Inc.

Applicant Address : F.5, NO.135, LANE 235, PAO CHIAO RD., HSIN-TIEN CITY, TAIPEI,
TAIWAN, R.O.C.

Manufacturer : AAEON Technology Inc.

Manufacturer Address : F.5, NO.135, LANE 235, PAO CHIAO RD., HSIN-TIEN CITY,
TAIPEI, TAIWAN, R.O.C.

Sample : Medical PC

Model / Brand Name : ONYX-172X / — (PS X=0~9 , or non)

Dimensions : 383 mm (L) x338 mm (W) x70 mm (D)

Standard : IEC 60529 (2001-02)

Sampling procedure : Sent by applicant.

Sampling date : June 3, 2009

Testing date : June 10, 2009

Issue Date : August 10, 2009

Testing engineer : Cheng, Sheng-Hsing

Result of test : The enclosure of Medical PC (Model : ONYX-172X (PS X=0~9 , or non)
) is satisfactory for IPX1, refer to page 2 for the test result.

Signatory of the report:

Sen-Li Lai



Result of Test

1.Result of test: The enclosure of Medical PC is satisfactory for IPX1.

2.Description of test:

Test items	Standards of test	Result of test
Degrees of protection against water. Second characteristic numeral 1.	IEC60529 (2001-02) sec.14.1 sec.14.2.1 sec.14.3	The protection is satisfactory.





財團法人

台灣大電力研究試驗中心
Taiwan Electric Research & Testing Center


Lighting Laboratory
Report No. : IPT20090089 -011

Pictures of sample:

View(1)



View(2)





View(3)

(Sealed with silicon gel)

