





1. <i>Test item list</i> -----	2
2. <i>Random Vibration Operation Test</i> -----	3

## Test Configuration:

Num	Item	Spec
1.	<b>Low Noise Medical Station:</b>	Onyx-217
	1.LCD	TFT LCD.17".CPT.CLAA170EA07.4 LAMP
	2.Power Adapter	EDAC EA1050A-120
	3. Inverter	DC TO AC.FOR17" TFT LCD (4LAMPS).QF132V1.16
	4. A/D Board	BT-R08LDNQ REV:05
	5. USB Transfer Board	T040 REV:A02
	6. USB Board	1907YC0301 REV:A1.0
	7. CD-ROM Transfer Board	1907T04101 REV:A0.2
	8. Smart card Board	MR0103
	9. Card Reader Board	GS-2004-CR18801 V1.1
	10. CD-ROM	TEAC DV-28SL
2.	<b>Test System</b>	Onyx-153D (PCM-8200 Rev: A1.0 / 512MB / BIOS:1.0)



**Test Date:** 02-22-2006

**Test Product:** Onyx-217

**Test Site:** AAEON QA Internal Lab.

**Performed By:** Rex Chang


**Test Standard:** Reference IEC68-2-64 Testing procedures  
Test Fh: Vibration boardband random test

**Test Equipment:**

Vibration Simulator System  
KING DESIGN Co. LTD.  
Model: KD 9363-EM-600F2K-40N20  
Serial Number: UU110099090  
Date of Calibration: 10/20/2005

**Test Condition:**

1. Operation
2. Test Acceleration: 1.5 G Random
3. Test Frequency: 5-500Hz
4. Test Axis: X, Y, Z axes
5. Test Time: 60min each axis
6. Test Software: Windows XP / Run one Microsoft media player simultaneously.
7. Test Vibration Curve:



**Test Result:**

The system structure doesn't deformation; Function is passed during system test.